

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Certificate of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

has been assessed and accredited in accordance with the Standard

ISO/IEC 17025:2017

“General Requirements for the Competence of Testing and Calibration Laboratories”

In the field of

Testing

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued compliance to the above standard & any other requirements specified by QAI.



QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

A handwritten signature in black ink, appearing to read 'Bhupendra'.

Dr. Bhupendra Kumar Rana
Chief Executive Officer

A handwritten signature in black ink, appearing to read 'Vikram'.

Prof. Vikram Kumar
Chair, CIA



Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
1.	Electrocardiograph	General requirements	IS 13450 (PART-2/ SEC 25): 2018 (Cl. No. 201.4) /IEC/EN 60601-2-25:2011
2.		General requirements for testing ME Equipment	IS 13450 (PART 2/SEC 25):2018 (Cl. No. 201.5)/IEC/EN 60601-2-25:2011
3.		Classification of ME Equipment and ME Systems	IS 13450 (PART 2/SEC 25):2018 (Cl. No. 201.6) /IEC/EN 60601-2-25:2011
4.	Automatically-controlled brachytherapy after loading equipment	ME Equipment identification, marking and documents	IS 13450 (Part-2/Sec-17: 2018 (Cl. No. 201.7)/ IEC/EN 60601-2-17: 2013
5.		Protection against electrical hazards from ME Equipment.	IS 13450 (Part-2/Sec-17: 2018(Cl. No. 201.8) /IEC/EN 60601-2-17: 2013
6.		Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450 (Part-2/Sec-17: 2018 (Cl. No. 201.9) /IEC/EN 60601-2-17: 2013
7.		Protection against excessive temperatures and other Hazards	IS 13450 (Part-2/Sec-17: 2018) (Cl. No. 201.11)/IEC/EN 60601-2-17: 2013
8.	Ultrasonic physiotherapy equipment	General Requirements for Test	IS 13450 (Part-2/ Sec-5):2018 (Cl. no. 4) /IEC/EN 60601-2-5:2009
9.		Classification	IS 13450 (Part 2 / Sec 5): 2018 (Cl. no. 5)/IEC/EN 60601-2-5:2009
10.		Identification, Marking and Documents Parts	IS 13450 (Part-2/ Sec-5):2018 (Cl. no.6)/IEC/EN 60601-2-5:2009

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



1 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
11.	Ultrasonic physiotherapy equipment	Power Input	IS 13450 (Part 2: Sec 5);2018 (Cl. no. 7)/IEC/EN 60601-2-5:2009
12.		General	IS 13450 (Part 2: Sec 5): 2018 (Cl. no. 13)/IEC/EN 60601-2-5:2009
13.		Mechanical strength	IS 13450 (Part 2: Sec 5): 2018 (Cl. no. 21)/IEC/EN 60601-2-5:2009
14.		Excessive Temperatures	IS 13450 (Part 2: Sec 5): 2018 (Cl. no. 42)/IEC/EN 60601-2-5:2009
15.		Components and general assembly	IS 13450 (Part 2: Sec 5): 2018 (Cl. no. 56)/IEC/EN 60601-2-5:2009
16.	Radiotherapy Simulators	Classification	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 5)/IEC/EN 60601-2-29:2009
17.		Identification, marking and documents	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 6)/IEC/EN 60601-2-29:2008
18.		Environmental conditions	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 10)/IEC/EN 60601-2-29:2008
19.		Enclosures And Protective Covers	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 16)/IEC/EN 60601-2-29:2008
20.		Protective earthing, functional earthing and potential equalization	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 18)/IEC/EN 60601-2-29:2008

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



2 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
21.		Moving parts	IS 13450 (Part 2: Sec 29): 2018 (Cl. no. 22)/IEC/EN 60601-2-29:2008
22.	Nerve and muscle stimulators	Classification of ME Equipment and ME Systems	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.6)/ IEC/EN 60601-2-10:2012 (withdrawn standard)
23.		ME Equipment identification, marking and documents	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201. 7)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)
24.		Protection against electrical hazards from ME Equipment	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.8)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)
25.		Protection against Mechanical L Hazards of ME Equipment and ME Systems	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.9)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)
26.		Protection against excessive temperatures and other hazards	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.11)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)
27.		Construction of ME equipment	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.15)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



3 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
28.		ME Systems	IS 13450: (Part 2: Sec 10): 2018 (Cl. no. 201.16)/ IEC/EN 60601-2-10: 2012 (withdrawn standard)
29.	Infant Incubators	General requirements	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.4)/IEC/EN 60601-2-19:2020
30.		General requirements for testing ME Equipment	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.5)/ IEC/EN 60601-2-19:2020
31.		Classification of ME Equipment and ME Systems	IS 13450 : (Part 2/sec 19) 2023 (Cl. No. 201.6) /IEC/EN 60601-2-19:2020
32.		ME Equipment identification, marking and documents	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.7) /IEC/EN 60601-2-19:2020
33.		Protection against electrical Hazards from ME Equipment	IS 13450; (Part 2/sec 19) 2023 (Cl. No. 201.8) / IEC/EN 60601-2-19:2020
34.		Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.9) /IEC/EN 60601-2-19:2020
35.		Protection against excessive temperatures and other Hazards	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.11) / IEC/EN 60601-2-19:2020
36.		Construction of ME Equipment	IS 13450: (Part 2/sec 19) 2023 (Cl. No. 201.15) / IEC/EN 60601

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



4 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
37.		ME Systems	IS 13450:(Part 2/sec 19) 2023 Cl. No. 201.16) / IEC/EN 60601
38.	Infant Transport Incubators	General requirements	IS 13450: (Part 2/sec 20) 2023 (Cl. No. 201.4) /IEC/EN 60601-2-20:2020
39.		General requirements for testing ME Equipment	IS 13450: (Part 2/sec 20) 2023 (Cl. No. 201.5)/ IEC/EN 60601-2-20:2020
40.		Classification of ME Equipment and ME Systems	IS 13450 :(Part 2/sec 20) 2023 (Cl. No. 201.6)/IEC/EN 60601-2-20:2020
41.		ME Equipment identification, marking and documents	IS 13450: (Part 2/sec 20) 2023 (Cl. No. 201.7)/IEC/EN 60601-2-20:2020
42.		Protection against electrical Hazards from ME Equipment	IS 13450; (Part 2/sec 20) 2023 (Cl. No. 201.8) / IEC/EN 60601-2-20:2020
43.		Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450: (Part 2/sec 20) 2023 (Cl. No.201.9) /IEC/EN 60601-2-20:2020
44.		Construction of ME Equipment	IS 13450 ;(Part 2/sec 20) 2023 (Cl. No. 201.15) / IEC/EN 60601-2-20:2020
45.		ME Systems	IS 13450:(Part 2/sec 20) 2023 (Cl. No. 201.16) / IEC/EN 60601:2020
46.	Electrocardiograph	General requirements	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.4)/IEC 60601-2-25: 2011

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



5 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
47.	Electrocardiograph	General requirements for testing ME Equipment	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.5)/IEC 60601-2-25: 2011
48.		Classification of ME Equipment and ME Systems	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.6)/IEC 60601-2-25: 2011
49.		ME Equipment identification, marking and documents	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.7)/IEC 60601-2-25: 2011
50.		Protection against Electrical equipment hazards from ME Equipment	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.8)/IEC 60601-2-25: 2011
51.		Protection against Electrical equipment hazards from ME Equipment.	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.9)/IEC 60601-2-25: 2011
52.		Protection Against Excessive temp. and other hazards.	IS 13450: (Part 2: Sec 25): 2018 (Cl. No. 201.11)/IEC 60601-2-25: 2011
53.	Operating Tables	Protection against electrical Hazards from Operating Tables	IS 13450 (Part 2/Sec 46): 2020 (Cl. No 201.8)/IEC/EN 60601-2-46:2016
54.		Protection against Mechanical Hazards of Operating Tables and ME Systems	IS 13450 (Part 2/Sec 46): 2020 (Cl. No 201.9) /IEC/ EN 60601-2-46:2016
55.		Mechanical Hazards associated with moving parts	IS 13450 (Part 2/Sec 46): 2020 (Cl. No 201.9.2.3) /IEC/EN 60601-2-46:2016

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



6 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
56.	Gamma Beam Therapy Equipment	Mechanical Hazards associated with support systems	IS 13450 (Part 2/Sec 46): 2020 (Cl. No 201.9.8 IEC /EN 60601-2-46:2016
57.		Protection against excessive temperatures and other Hazards	IS 13450 (Part 2/Sec 46): 2020 (Cl. No 201.11)/ IEC/EN 60601-2-46:2016
58.		Protection against electric shock	IS 13450 (Part 2/Sec 11): 2018 (Cl. 201.6.2)/IEC/EN 60601-2-11: 2013
59.		Protection against harmful ingress of water or particulate matter	IS 13450 (Part 2/Sec 11): 2018(Cl. 201.6.3)/IEC/EN 60601-2-11: 2013
60.		Marking on the inside of ME Equipment or ME Equipment parts	IS 13450 (Part 2/Sec 11): 2018 (Cl. 201.7.3) /IEC/EN 60601-2-11: 2013
61.		Protection against electrical Hazards from ME Equipment	IS 13450 (Part 2/Sec 11): 2018 (Cl. 201.8) /IEC/EN 60601-2-11: 2013 (Excluding Cl. No. 201.8.7)
62.		Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450 (Part 2/Sec 11): 2018 (Cl. 201.9) IEC/EN 60601-2-11: 2013 (Excluding Cl. No. 209.2 and 209.7)
63.	Invasive blood pressure monitoring equipment	Protection against electric shock	IS 13450 (Part 2/Sec 34): 2019 (CL.201.6.2); IEC/EN 60601-2-34: 2011

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



7 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
64.	Invasive blood pressure monitoring equipment	Marking on the outside of ME Equipment or ME Equipment parts	IS 13450 (Part 2/Sec 34): 2019 (CL. 201.7.2); IEC/EN 60601-2-34: 2011
65.		Protection against electrical Hazards from ME Equipment	IS 13450 (Part 2/Sec 34): 2019 (CL. 201.8); IEC/EN 60601-2-34:2011 (Excluding Cl No. 201.8.5.5.1 201.8.7.4.7)
66.		Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450 (Part 2/Sec 34): 2018 (CL. 201.9) IEC/EN 60601-2-34: 2013 (Excluding Cl No. 201.9.7)
67.		Protection against excessive temperatures and other Hazards	IS 13450 (Part 2/Sec 34): 2019 (CL. 201.11) IEC/EN 60601-2-34: 2011
68.		Drop test	IS 13450 (Part 2/Sec 34): 2019 (CL. 201.15.3.4.101)/ IEC/EN 60601-2-34:2011
69.		ME Equipment components and general assembly	IS 13450 (Part 2/Sec 34): 2019 (CL. 201.15.4)/ IEC/EN 60601-2-34:2011
70.	Infant Radiant Warmers	Protection against electrical Hazards from ME Equipment	IS 13450 (Part 2 /Sec 21): 2023 (CL. 201.8) /IEC/EN 60601-2-21: 2020

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



8 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
71.	Infant Radiant Warmers	Protection against Mechanical Hazards of ME Equipment and ME Systems	IS 13450 (Part 2 /Sec 21): 2023 (CL. 201.9) IEC/EN 60601-2-21: 2020 (Excluding Cl No. 201.9.6.2)
72.		Protection against excessive temperatures and other Hazards	IS 13450 (Part 2 /Sec 21): 2023 (CL. 201.11) IEC/EN 60601-2-21: 2020 (Excluding Cl 201.11.6.3)
73.		Construction of connectors	IS 13450 (Part 2/Sec 21): 2023 (CL.201.15.4.1) IEC/EN 60601-2-21: 2020(Excluding Cl No. 201.15.4.2.1)
74.	Surgical luminaires and luminaires for diagnosis	Protection against electric shock	IS 13450 (Part 2/Sec 41): 2020 (CL. 201.6.2) IEC/EN60601-2 -41: 2009
75.		Protection against electrical Hazards from ME Equipment	IS 13450 (Part 2/Sec 41): 2020 (CL.201.8) IEC/EN60601-2 -41: 2009
76.		Protection against mechanical hazards of ME Equipment and ME Systems	IS 13450 (Part 2/Sec 41): 2020 (CL.201.9) IEC/EN60601-2 -41: 2009
77.		Detachable Handle	IS13450 (Part 2/Sec 41): 2020 (CL. 201.9.2.101)/IEC/EN60601-2 -41: 2009
78.		Protection against excessive temperatures and other Hazards	IS 13450 (Part 2/Sec 41): 2020 (CL. 201.11) IEC/EN60601-2 -41: 2009 Excluding Cl. No. 201.11.8.101)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



9 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
79.		Mechanical strength	IS 13450 (Part 2/Sec 41): 2020 (CL.201.15.3) IEC/EN60601-2-41: 2009
80.	Multifunction Patient Monitoring Equipment	Identification, Marking and documents	IS 13450 (Part 2/Sec 49): 2009 (Cl. NO. 5)IEC/EN 60601-2-49 : 2001 (withdrawn standard)
81.		Requirement related to classification	IS 13450 (Part 2/Sec 49): 2009 (Cl. NO. 14)IEC/EN 60601-2-49 : 2001 (withdrawn standard)
82.		Dielectric strength	IS 13450 (Part 2/Sec 49): 2009 (Cl. NO. 20) IEC/EN 60601-2-49: 2001(withdrawn standard)
83.	Electro-Cardiographic Monitoring Equipment	Protection against electric shock	IS 13450 (Part 2/Sec 27): 2018(CL. 201.6.2) IEC/EN 60601-2-27:2011
84.		Protection against electrical Hazards from ME Equipment	IS 13450 (Part 2/Sec 27): 2018 (CL. 201.8); IEC/EN 60601-2-27: 2011 (Excluding Cl . No. 201.8.5.5)
85.		Patient leads	IS 13450 (Part 2/Sec 27): 2018 (CL. 201.8.5.2.3) IEC/EN 60601-2-27: 2011
86.		Ingress of water or particulate matter into ME Equipment and ME Systems	IS 13450 (Part 2/Sec 27): 2018 (CL. 201.11.6.5) IEC/EN 60601-2-27: 2011

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



10 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
87.		Drop test	IS 13450 (Part 2/Sec 27): 2018 (CL. 201.15.3.4) IEC/EN 60601-2-27: 2011
88.	General Equipments	Power Supply	IS 13450 (Part-1): 2018 (CL.4.10) / IEC 60601-1:2012 (Withdrawn standard)
89.		Power Input	IS 13450 (Part-1): 2018 (CL.4.11) / IEC 60601-1:2012 (Withdrawn standard)
90.		Legibility of marking	IS 13450 (Part-1): 2018 (CL.7.1.2) / IEC 60601-1:2012 (Withdrawn standard)
91.		Durability Markings	IS 13450 (Part-1): 2018 (CL.7.1.3) / IEC 60601-1:2012 (Withdrawn standard)
92.		Temperature of supply terminal	IS 13450 (Part-1): 2018 (CL.7.3.8) / IEC 60601-1:2012 (Withdrawn standard)
93.		Safety Sign	IS 13450 (Part-1): 2018 (CL.7.5) / IEC 60601-1:2012 (Withdrawn standard)
94.		Symbol for control and Performance	IS 13450 (Part-1): 2018 (CL.7.6.3) / IEC 60601-1:2012 (Withdrawn standard)
95.		Dielectric Strength	IS 13450 (Part-1): 2018 (CL. 8.8.3) / IEC 60601-1:2012 (Withdrawn standard)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



11 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
96.	General Equipments	Insulation strength and resistance to heat	IS 13450 (Part-1): 2018 (CL. 8.8.4.1)/ IEC 60601-1:2012 (Withdrawn standard)
97.		Component and wiring	IS 13450 (Part-1): 2018 (CL. 8.10) / IEC 60601-1:2012 (Withdrawn standard)
98.		Mechanical Strength	IS 13450 (part 1): 2018 (CL.15.3)/ IEC 60601-1: 2012 (Withdrawn standard)
99.		Short –Circuit test	IS 13450 (Part-1): 2018 (CL.15.5.1.2)/ IEC 60601-1:2012 (Withdrawn standard)
100.		Overload test	IS 13450 (Part-1): 2018(CL.15.5.1.3)/IEC 60601-1:2012 (Withdrawn standard)
101.		ME System Connection and Wiring	IS 13450 (Part-1): 2018 (CL.16.9 / IEC 60601-1:2012 (Withdrawn standard)
102.	General Medical Electrical System	Humidity preconditioning treatment	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 , Cl.5.7 (Withdrawn standard)
103.		ME Equipment identification, marking and documents	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 , Cl.7.0 (Withdrawn standard)
104.		Creepage distances and air clearances	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.8.9 (Withdrawn standard)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



12 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
105.	General Medical Electrical System	Material groups classification Tracking Test	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.8.9.1.7 (Withdrawn standard)
106.		Cord anchorage	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.8.11.3.5
107.		Cord guards	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.8.11.3.6 (Withdrawn standard)
108.		Instability hazards	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.9.4 (Withdrawn standard) (Excluding Cl No. 9.4.2.4.2 & Cl No. 9.4.2.4.3)
109.		Excessive temperatures	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 Cl.11.1 (Withdrawn standard)
110.		Constructional requirements for fire enclosures	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 Cl.11.3 (Withdrawn standard)
111.		Temperature and Overload Control Devices	IS 13450 Part 1 :2018 Cl 15.4.2 /IEC 60601 -1 :2012 (Withdrawn standard)
112.		Actuating parts of controls	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 Cl.15.4.6 (Withdrawn standard)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



13 of 14

Quality And Accreditation Institute

Centre for International Accreditation

(formerly Centre for Laboratory Accreditation)



Scope of Accreditation

Planet Electro Labs Private Limited

A-74/1, Wazirpur Industrial Area

New Delhi-110052, India

QAI/CIA/TL/2025/0130

Valid from: 14 July 2025

Valid until: 13 July 2027

Accreditation Standard: ISO/IEC 17025:2017

Electronics Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Medical Electrical Equipment		
113.	General Medical Electrical system	Cord-connected hand-held and foot-operated control devices	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 Cl.15.4.7 (Withdrawn Standard)
114.		Transformers	IS 13450 (Part -1): 2018 / IEC 60601-1:2012 Cl.15.5.1.1 (Withdrawn standard)
115.		Leakage currents	IS 13450 (Part -1): 2018 / IEC 60601-1:2012, Cl.16.6 (Withdrawn standard) (Exclusion)
116.		Protection against excessive temperature and Other Hazards	IS/ISO 11197: 2019; Cl 201.11
117.		Protection Against Electric Shock	IS 17512:2021 (CL. 6.1)

